

## Section 1 - Evaluation of ACR PET Phantom


| PET Acquisition Parameters |                 | CT Acquisition Parameters |  | Reconstruction Techniques |       |
|----------------------------|-----------------|---------------------------|--|---------------------------|-------|
| Beds                       |                 | Topo mAs                  |  | Reconstruction            |       |
| Time/Bed                   |                 | Topo kVp                  |  | Iterations                |       |
| Matrix                     |                 | mAs                       |  | Subsets                   |       |
| Zoom                       | 1               | kVp                       |  | Filter                    |       |
| Isotope                    | <sup>18</sup> F | No. Slices                |  | Setting                   |       |
| Activity                   |                 | Slice Thickness           |  | Slice Thickness           | 10 mm |

**Note** Images should be acquired with whole body protocol with the same settings that are used for routine clinical studies. The entire phantom should be reconstructed with the same protocol used for whole body scans including pre- and post- reconstruction filters. Slices should be 1 cm thick.

| ACR PET Phantom Dilution |      |      |                 |
|--------------------------|------|------|-----------------|
|                          | Dose | Time | Dose Ratios     |
| Patient Dose             |      |      |                 |
| FDG Dose (A) mCi         |      |      | FDG Doses: B/A  |
| FDG Dose (B) mCi         |      |      |                 |
| Test Dose #1 μCi         |      |      | Test Doses: 1/2 |
| Test Dose #2 μCi         |      |      |                 |
| Actual Start Time        |      |      |                 |

Transaxial Images Saved In 10mm Slices Saved

## Section 2 - SUV Evaluation



| Measured SUV's |     |      |     |
|----------------|-----|------|-----|
| Cylinders      | Min | Mean | Max |
| Bone           |     |      |     |
| Air            |     |      |     |
| Water          |     |      |     |
| 25 mm          |     |      |     |
| 16 mm          |     |      |     |
| 12 mm          |     |      |     |
| 8 mm           |     |      |     |
| Background     |     |      |     |

SUV Image Saved With Mean, Min, and Max SUV Values

## Section 3 - PET/CT Gantry Alignment

| Notes  | Pass / Fail |
|--|-------------|
| Does PET/CT Gantry demonstrate adequate alignment? |             |

Fused Transaxial and Axial Images Saved

## Section 4 - Count Rate

| Phantom Activity | Total Count Rate | Mode | Pass / Fail |
|------------------|------------------|------|-------------|
|                  |                  | 2D   |             |
|                  |                  | 3D   |             |

| Notes                                 |
|---------------------------------------|
| Include all info for Siemens scanners |

## Section 5 - Video Display

| SMPTE Steps | Measured Luminance (Cd/m <sup>2</sup> ) |
|-------------|---|
| 0%          |   |
| 10%         |   |
| 20%         |   |
| 30%         |   |
| 40%         |   |
| 50%         |   |
| 60%         |   |
| 70%         |   |
| 80%         |   |
| 90%         |   |
| 100%        |   |

| Visual Assessment     | Pass/Fail |
|-----------------------|-----------|
| 95% Patch Visible     |           |
| 5% Patch Visible      |           |
| Distortion Acceptable |           |
| Tearing or Blurring   |           |
| Focus Acceptable      |           |

| Uniformity |   |
|------------|---|
| Pattern    | Measured Luminance (Cd/m <sup>2</sup> ) |
| Top        |   |
| Bottom     |   |
| Left       |   |
| Right      |   |
| Center     |   |

## Section 6 - System Interlocks

|                           |   |  |
|---------------------------|---|--|
| Gantry Housing            | The gantry housing is intact and adequately secured?  |  |
| Cables                    | Cables are secure and void of fraying or exposed wiring?  |  |
| Switches                  | All switches are functioning correctly? No errors were noted when powering the camera on/off?               |  |
| Gantry Controls & Motions | Gantry control and motions functioning as designed?   |  |
| Table                     | The table is damage free and is capable of supporting patient loads? Table motion adequate for this system? |  |
| Emergency Stop            | Emergency stop switches functioning as designed?  |  |

## Section 7 - Dose Calibrator

| Dose Calibrator |           |           |
|-----------------|-----------|-----------|
| Test            | Test Date | Pass/Fail |
| Constancy       |           |           |
| Linearity*      |           |           |
| Accuracy*       |           |           |

|                             | Date | Pass/Fail |
|-----------------------------|------|-----------|
| Most Recent NRC Inspection* |      |           |

\* must be submitted to ACR

**Section 8 - Quality Control**

| PET QC                   | Frequency | Last Performed |
|--------------------------|-----------|----------------|
| DQA (Daily QA)           | Daily     |                |
| Normalization            | Quarterly |                |
| Well Counter Calibration | Quarterly |                |
|                          |           |                |
|                          |           |                |
| CT QC                    | Frequency | Last Performed |
|                          |           |                |
|                          |           |                |
|                          |           |                |
|                          |           |                |
|                          |           |                |
|                          |           |                |
|                          |           |                |
|                          |           |                |
| Display QC               | Frequency | Last Performed |
| SMPTE Evaluation         | Monthly   |                |

| Comments   |  |
|--|--|
| <ul style="list-style-type: none"> <li>•</li> </ul> |  |